




Digimatic Calipers

IP67 ABS Digimatic Coolant Proof Calipers

ABS Digimatic Calipers



Catalog No. E12047(2)





ABS Digimatic Calipers

MeasurLink® ENABLED

ABSOLUTE™







Products equipped with the measurement data output function can be connected to the measurement data network system MeasurLink[®].

Note: MeasurLink® is a registered trademark of Mitutoyo Corporation in Japan and Mitutoyo America Corporation in the United States

ABS (Absolute) scale makes origin-setting at every power-on unnecessary, and the allowable slider speed is unlimited with this type of scale.



High quality guide surface finish for smooth slider movement



ABS Digimatic Caliper ex. **500-151-30**

4

Mitutoyo

Common specifications

(IP67 ABS Digimatic Coolant Proof Calipers, ABS Digimatic Calipers)

- Resolution: 0.01 mm
- Power supply: Silver oxide button cell battery SR44 (938882), 1 pc. supplied for testing purposes only. Replace with a fresh cell as soon as the caliper is put into use.
- Battery life: Approx. 5 years under normal use
- Max. response speed: Unlimited

Common functions

(IP67 ABS Digimatic Coolant Proof Calipers, ABS Digimatic Calipers)

Absolute measurement:

After power is turned ON, measurement can be started without zerosetting if origin-setting was previously performed. The Absolute origin position can be changed by the ORIGIN button.

ORIGIN button:

Sets the Absolute origin at the current slider position when pressed.

Switches between Absolute and Incremental measurement modes when pressed.

IP67 ABS Digimatic Coolant Proof Calipers

Specifications

Order No.	Measuring range (mm)	Maximum permissible error (mm)*		Mass	Measurement	Thumb roller	Remarks
		<i>Е</i> мре	Ѕмре	(g)	data output port	THUITID TOILEI	Neillaiks
500-702-20	0 - 150	±0.02	±0.04	168	_	✓	_
500-703-20	0 - 200			198	_	✓	_
500-712-20	0 - 150			168	✓	✓	_
500-713-20	0 - 200			198	✓	✓	_
500-714-20	0 - 300	±0.03	±0.05	350	✓	✓	_
500-721-20	0 - 150	±0.02	±0.04	168	✓	✓	Carbide-tipped jaws for outside measurement
500-722-20	0 - 200			198	✓	✓	Carbide-tipped jaws for outside measurement
500-723-20	0 - 150			168	✓	✓	Carbide-tipped jaws for outside & inside measurement
500-724-20	0 - 200			198	✓	✓	Carbide-tipped jaws for outside & inside measurement
500-719-20	0 - 150			168	√	✓	Rod depth bar ø1.9 mm


^{*} Partial Surface Contact Error, EMPE and Shift Error, SMPE are terms (notations) used in ISO 13385-1:2019.

Inch/Metric

Order No.	Measuring range (in)	Maximum permissible error*		Mass	 Measurement	Thursh veller	Remarks
Order No.		Емре	Ѕмре	(g)	data output port	Thumb roller	REITIAIKS
500-752-20	0 - 6		±0.002 in/ ±0.04 mm	168	_	✓	_
500-753-20	0 - 8	±0.001 in/ ±0.02 mm		198	_	✓	_
500-762-20	0 - 6			168	✓	✓	_
500-763-20	0 - 8			198	✓	✓	_
500-764-20	0 - 12	±0.0015 in/ ±0.03 mm	±0.0025 in/ ±0.05 mm	350	✓	✓	_
500-735-20	0 - 6	±0.001 in/ ±0.02 mm	±0.002 in/ ±0.04 mm	168	✓	✓	Carbide-tipped jaws for outside measurement
500-736-20	0 - 8			198	✓	✓	Carbide-tipped jaws for outside measurement
500-737-20	0 - 6			168	✓	✓	Carbide-tipped jaws for outside & inside measurement
500-738-20	0 - 8			198	✓	✓	Carbide-tipped jaws for outside & inside measurement
500-769-20	0 - 6			168	✓	✓	Rod depth bar ø0.075 inch

^{*} Partial Surface Contact Error, EMPE and Shift Error, SMPE are terms (notations) used in ISO 13385-1:2019.

DIMENSIONS



Low-voltage alert:

If the battery voltage becomes low, a "B" appears in the display to alert the user before measurement is no longer possible. A battery change advisory alert precedes this alert.

By using the connecting cable (option), measurement data can be output.

ABS Digimatic Calipers

Specifications


Metric

Order No.	Measuring range (mm)	Maximum permissible error (mm)*2		Mass	Measurement	Thumb roller	Remarks
		Емре	Ѕмре	(g)	data output port	Thumb folier	Remarks
500-150-30	0 - 100	ĺ		143	✓	✓	Rod depth bar ø1.9 mm
500-180-30*1	0 - 100			143	_	_	Rod depth bar ø1.9 mm
500-151-30		±0.02	±0.04	168	✓	✓	_
500-154-30					✓	✓	Carbide-tipped jaws for outside measurement
500-155-30	0 - 150				✓	✓	Carbide-tipped jaws for outside & inside measurement
500-158-30					✓	✓	Rod depth bar ø1.9 mm
500-181-30*1					_	_	_
500-152-30				198	✓	✓	_
500-156-30	0 - 200				✓	✓	Carbide-tipped jaws for outside measurement
500-157-30					✓	✓	Carbide-tipped jaws for outside & inside measurement
500-182-30*1						_	_
500-153-30	0 - 300	±0.03	±0.05	350	✓	✓	_

Inch/Metric

Order No.	Measuring range (in)	Maximum permissible error* ²		Mass	Measurement	Thomas vallas	Remarks
Order No.		Емре	Smpe	(g)	data output port	Thumb roller	Remarks
500-170-30	0 - 4			143	✓	✓	Rod depth bar ø0.075 inch
500-171-30		±0.001 in/ ±0.02 mm	±0.002 in/ ±0.04 mm	168	✓	✓	_
500-174-30]				✓	✓	Carbide-tipped jaws for outside measurement
500-175-30	0 - 6				✓	✓	Carbide-tipped jaws for outside & inside measurement
500-178-30]				✓	✓	Rod depth bar ø0.075 inch
500-191-30					_		<u> </u>
500-172-30					✓	✓	_
500-176-30	0 - 8				✓	✓	Carbide-tipped jaws for outside measurement
500-177-30					✓	✓	Carbide-tipped jaws for outside & inside measurement
500-173-30	0 - 12	±0.0015 in/ ±0.03 mm	±0.0025 in/ ±0.05 mm	350	✓	✓	_
500-195-30*1	0 - 4	±0.001 in/ ±0.02 mm	±0.002 in/ ±0.04 mm	143	_	✓	Rod depth bar ø0.075 inch
500-196-30*1	0 - 6			168	_	✓	_
500-197-30*1	0 - 8			198		✓	_


DIMENSIONS



^{*1} Without SPC data output
*2 Partial Surface Contact Error, EMPE and Shift Error, SMPE are terms (notations) used in ISO 13385-1:2019.

^{*1} Without SPC data output
*2 Partial Surface Contact Error, EMPE and Shift Error, SMPE are terms (notations) used in ISO 13385-1:2019.

Centralized data management



Optional cable accessories

	Product	Order No.
	Connecting coble	959149 (1 m)
ABS Digimatic Calipers	Connecting cable	959150 (2 m)
	Hold unit	959143
IP67 ABS Digimatic	Connecting cable	05CZA624 (1 m)
Coolant Proof Calipers	Connecting cable	05CZA625 (2 m)

Combining optional accessories enables not only wireless measurement data recording but also advanced statistical processing management.



Find additional product literature and our product catalogue

https://www.mitutoyo.co.jp/global.html

Our products are classified as regulated items under Japanese Foreign Exchange and Foreign Trade Law. Please consult us in advance if you wish to export our products to any other country.

If the purchased product is exported, even though it is not a regulated item (Catch-All controls item), the customer service available for that product may be affected. If you have any questions, please consult your local Mitutoyo sales office.

Note: Product illustrations are without obligation. Product descriptions, in particular any and all technical specifications, are only binding when explicitly agreed upon.

MITUTOYO and MiCAT are either registered trademarks or trademarks of Mitutoyo Corp. in Japan and/or other countries/regions.

Other product, company and brand names mentioned herein are for identification purposes only and may be the trademarks of their respective holders.

102 2110(2)g(DI)CM-(JP2), Printed in Japan



Mitutoyo Corporation

20-1, Sakado 1-Chome,

Takatsu-ku, Kawasaki-shi,

Kanagawa 213-8533, Japan

T +81 (0) 44 813-8230

F +81 (0) 44 813-8231

https://www.mitutoyo.co.jp